

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5196

Name: RITTER 1  
Sterilizer ID: V2178198  
12/17/2024 09:16 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.3 F 0.0  
2:00 80.0 F 0.0  
4:00 94.8 F 0.0  
6:00 128.7 F 0.0  
8:00 163.2 F 0.1  
10:00 187.8 F 0.2  
12:00 205.6 F 0.8  
14:00 218.4 F 4.4  
16:00 226.6 F 7.1  
18:00 235.6 F 10.7  
20:00 234.0 F 7.9  
22:00 243.6 F 12.4  
24:00 240.6 F 10.6  
26:00 237.5 F 9.1  
28:00 252.3 F 16.2  
30:00 264.4 F 23.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 272.7 F 28.6  
2:00 271.7 F 28.0  
3:00 272.8 F 28.6  
4:00 271.6 F 27.9

Min 271.6 F 27.8  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.2 F 0.0  
5:00 199.9 F 0.0  
10:00 211.6 F 0.0  
15:00 216.7 F 0.0  
20:00 216.9 F 0.0  
25:00 220.3 F 0.0  
30:00 216.1 F 0.0  
Drying Complete

FILLING: 0:40  
HEATING: 31:56  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:09:55  
CYCLE COMPLETE

12/17/2024 10:25 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_