
Midmark M11 - v1.0.5
Total Cycles: 891

Name: RITTER2
Sterilizer ID: V2313228
03/21/2022 02:22 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.5 F 0.0
2:00 117.8 F 0.0
4:00 166.0 F 0.0
6:00 181.0 F 0.3
8:00 192.4 F 2.3
10:00 211.0 F 7.2
12:00 226.5 F 12.2
14:00 232.2 F 8.5
16:00 241.0 F 11.4
18:00 244.6 F 12.9
20:00 243.7 F 12.2
22:00 257.3 F 19.5
24:00 269.4 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.9 F 28.5
2:00 272.5 F 28.9
3:00 273.0 F 29.3

Min 271.6 F 28.4
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 216.6 F 0.0
10:00 219.3 F 0.0
Dry Cycle Aborted

FILLING: 1:23
HEATING: 24:55
STERILIZING: 3:00
VENTING: 2:49
DRYING: 12:53
TOTAL CYCLE: 00:45:00

03/21/2022 03:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____