
Midmark M11 - v1.0.5
Total Cycles: 2645

Name: RITTER2
Sterilizer ID: V2313228
07/14/2023 08:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.3 F 0.0
2:00 75.1 F 0.0
4:00 101.9 F 0.0
6:00 101.2 F 0.0
8:00 154.2 F 0.8
10:00 176.7 F 3.9
12:00 192.1 F 7.4
14:00 211.2 F 11.1
16:00 218.2 F 5.3
18:00 229.6 F 8.7
20:00 238.2 F 12.9
22:00 236.1 F 9.8
24:00 239.4 F 9.6
26:00 246.4 F 13.8
28:00 257.5 F 19.9
30:00 267.7 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.6
2:00 271.5 F 28.8
3:00 272.6 F 29.7
4:00 272.1 F 29.1

Min 271.4 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 219.0 F 0.0
10:00 219.1 F 0.0
15:00 213.7 F 0.0
20:00 214.3 F 0.0
25:00 215.5 F 0.0
30:00 214.6 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:26
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

07/14/2023 09:32 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____