
Midmark M11 - v1.0.5
Total Cycles: 3184

Name: RITTER2
Sterilizer ID: V2313228
12/05/2023 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 164.4 F 1.5
4:00 181.8 F 5.5
6:00 203.8 F 9.8
8:00 216.7 F 9.1
10:00 226.3 F 9.7
12:00 230.8 F 9.0
14:00 240.8 F 12.9
16:00 241.5 F 11.8
18:00 254.0 F 18.1
20:00 265.0 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.4 F 29.9
2:00 271.9 F 29.4
3:00 271.1 F 29.2
4:00 272.0 F 30.0

Min 271.1 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 208.5 F 0.0
10:00 210.7 F 0.0
15:00 203.5 F 0.0
20:00 201.4 F 0.0
25:00 201.0 F 0.0
30:00 200.3 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 21:60
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:00
CYCLE COMPLETE

12/05/2023 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____