
Midmark M11 - v1.0.0
Total Cycles: 2805

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2023 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.8 F 0.0
2:00 187.8 F 3.2
4:00 201.2 F 8.2
6:00 220.3 F 13.5
8:00 228.1 F 9.0
10:00 236.8 F 9.7
12:00 242.5 F 12.4
14:00 241.7 F 11.2
16:00 255.0 F 18.0
18:00 266.7 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 29.0
2:00 271.3 F 28.3
3:00 272.4 F 29.1
4:00 272.0 F 28.6

Min 271.3 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 228.0 F 0.0
10:00 229.0 F 0.0
15:00 238.9 F 0.0
20:00 228.6 F 0.0
25:00 224.3 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

01/23/2023 01:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____