
Midmark M11 - v1.0.0
Total Cycles: 2097

Name: RITTER 1
Sterilizer ID: V2178198
07/07/2022 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.4 F 0.0
2:00 94.8 F 0.0
4:00 108.5 F 0.2
6:00 131.8 F 2.0
8:00 164.6 F 4.4
10:00 203.8 F 8.2
12:00 212.0 F 12.6
14:00 222.9 F 6.3
16:00 231.6 F 10.5
18:00 233.6 F 7.6
20:00 241.2 F 11.7
22:00 237.5 F 9.1
24:00 250.5 F 15.3
26:00 261.3 F 21.4
28:00 271.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.8 F 28.6
2:00 271.8 F 28.5
3:00 272.1 F 28.6
4:00 272.2 F 28.7

Min 271.7 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 235.3 F 0.0
10:00 226.1 F 0.0
15:00 228.9 F 0.0
20:00 228.4 F 0.0
25:00 232.3 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:17
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:07:20
CYCLE COMPLETE

07/07/2022 10:39 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____