
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001618
16:33 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:02
STERILIZING 4:06
136.4 C 225kPa 4:06
136.3 C 225kPa 4:37
136.2 C 222kPa 5:07
136.2 C 224kPa 5:37
136.3 C 225kPa 6:07
136.3 C 225kPa 6:37
136.4 C 226kPa 7:07
136.4 C 224kPa 7:37
136.4 C 224kPa 7:37
VENTING 7:37
AIR DRYING 8:31
CYCLE COMPLETE 8:37

Digital Signature #
11E212913A1C91CF
