
Midmark M11 - v1.0.0
Total Cycles: 5201

Name: RITTER 1
Sterilizer ID: V2178198
12/18/2024 11:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.6 F 0.0
2:00 152.0 F 0.0
4:00 170.6 F 0.1
6:00 190.1 F 0.5
8:00 207.7 F 3.9
10:00 217.5 F 7.6
12:00 228.8 F 11.6
14:00 228.9 F 6.8
16:00 240.7 F 11.9
18:00 238.2 F 9.6
20:00 239.2 F 9.4
22:00 248.7 F 14.2
24:00 260.1 F 20.5
26:00 270.3 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.2 F 28.5
2:00 272.6 F 28.6
3:00 272.1 F 28.2
4:00 271.6 F 27.8

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 190.5 F 0.0
10:00 201.5 F 0.0
15:00 202.7 F 0.0
20:00 199.8 F 0.0
25:00 194.4 F 0.0
30:00 191.0 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 27:01
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:05:27
CYCLE COMPLETE

12/18/2024 12:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____