
Midmark M11 - v1.0.5
Total Cycles: 4103

Name: RITTER2
Sterilizer ID: V2313228
11/08/2024 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.9 F 0.0
2:00 206.5 F 3.5
4:00 210.4 F 10.2
6:00 223.2 F 6.4
8:00 238.8 F 12.9
10:00 240.7 F 11.6
12:00 241.3 F 11.3
14:00 256.3 F 19.1
16:00 269.1 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.3 F 29.1
2:00 272.8 F 29.6
3:00 271.7 F 28.8
4:00 272.7 F 29.2

Min 271.7 F 28.6
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 215.6 F 0.0
10:00 220.1 F 0.0
15:00 215.6 F 0.0
20:00 225.2 F 0.0
25:00 224.0 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 16:58
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:56:41
CYCLE COMPLETE

11/08/2024 02:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____