
Midmark M11 - v1.0.5
Total Cycles: 3519

Name: RITTER2
Sterilizer ID: V2313228
04/11/2024 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.2 F 0.0
2:00 121.4 F 1.3
4:00 150.6 F 4.4
6:00 175.4 F 7.7
8:00 193.9 F 12.0
10:00 220.1 F 5.7
12:00 227.8 F 10.2
14:00 234.3 F 8.3
16:00 241.1 F 11.8
18:00 238.3 F 10.0
20:00 250.9 F 16.2
22:00 262.1 F 22.6
24:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 271.4 F 29.7
2:00 271.6 F 29.8
3:00 271.5 F 29.7
4:00 271.2 F 29.6

Min 271.2 F 29.4
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 226.3 F 0.0
10:00 213.5 F 0.0
15:00 188.9 F 0.0
20:00 185.2 F 0.0
25:00 183.6 F 0.0
30:00 181.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 25:02
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:03:38
CYCLE COMPLETE

04/11/2024 04:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____