
Midmark M11 - v1.0.0
Total Cycles: 4903

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2024 04:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.2 F 0.0
2:00 187.8 F 0.6
4:00 214.6 F 4.9
6:00 218.2 F 9.8
8:00 230.3 F 7.9
10:00 237.3 F 10.7
12:00 236.4 F 8.2
14:00 241.7 F 10.6
16:00 247.5 F 13.8
18:00 259.6 F 20.4
20:00 270.3 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.6 F 29.0
2:00 272.0 F 28.4
3:00 271.3 F 28.2
4:00 272.5 F 29.0

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 217.1 F 0.0
10:00 221.9 F 0.0
15:00 235.2 F 0.0
20:00 238.0 F 0.0
25:00 232.7 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 20:54
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 00:59:58
CYCLE COMPLETE

09/12/2024 05:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____