
Midmark M11 - v1.0.0
Total Cycles: 5502

Name:
Sterilizer ID: V2178198
04/11/2025 05:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.6 F 0.0
2:00 167.9 F 3.4
4:00 176.8 F 7.5
6:00 200.3 F 11.9
8:00 212.6 F 4.8
10:00 225.6 F 9.5
12:00 234.7 F 9.3
14:00 239.2 F 11.7
16:00 238.4 F 9.9
18:00 251.7 F 16.5
20:00 263.3 F 23.2
22:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.0 F 28.7
2:00 272.1 F 28.9
3:00 272.4 F 29.2
4:00 271.9 F 29.0

Min 271.6 F 28.6
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 186.7 F 0.0
10:00 185.6 F 0.0
15:00 193.4 F 0.0
20:00 201.2 F 0.0
25:00 215.9 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:03
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:00:51
CYCLE COMPLETE

04/11/2025 06:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____