

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006424  
8:50 10/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:56  
PRESSURIZING 2:58  
STERILIZING 4:00  
136.5 C 222kPa 4:00  
136.0 C 222kPa 4:31  
136.4 C 223kPa 5:01  
136.3 C 223kPa 5:31  
136.4 C 224kPa 6:01  
136.3 C 218kPa 6:31  
136.4 C 222kPa 7:01  
136.1 C 219kPa 7:31  
136.1 C 219kPa 7:31  
VENTING 7:31  
AIR DRYING 8:30  
CYCLE COMPLETE 11:03

-----  
Digital Signature #  
FB5F535337867F99  
-----