
Midmark M11 - v1.0.0
Total Cycles: 1184

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2021 06:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 72.9 F 0.0
4:00 94.5 F 0.0
6:00 114.0 F 0.0
8:00 139.6 F 0.1
10:00 163.2 F 0.2
12:00 184.0 F 0.6
14:00 200.5 F 3.3
16:00 213.2 F 6.2
18:00 221.9 F 9.3
20:00 230.5 F 12.8
22:00 231.9 F 8.1
24:00 242.5 F 12.7
26:00 239.5 F 10.2
28:00 237.5 F 8.7
30:00 249.2 F 14.6
32:00 260.3 F 20.7
34:00 270.5 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 27.9
1:00 271.9 F 28.3
2:00 272.6 F 28.5
3:00 271.7 F 27.9
4:00 271.8 F 28.3

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 210.9 F 0.0
10:00 210.3 F 0.0
15:00 210.7 F 0.0
20:00 203.9 F 0.0
25:00 203.6 F 0.0
30:00 201.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 34:47
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:13:04
CYCLE COMPLETE

09/20/2021 08:11 AM
ACCEPTED: X
REJECTED:
Operator: U01