
Midmark M11 - v1.0.5
Total Cycles: 4256

Name: RITTER2
Sterilizer ID: V2313228
01/10/2025 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.1 F 0.0
2:00 162.2 F 0.1
4:00 181.6 F 1.6
6:00 199.6 F 5.4
8:00 214.9 F 9.6
10:00 227.6 F 9.5
12:00 234.6 F 10.2
14:00 235.4 F 9.0
16:00 243.5 F 12.1
18:00 246.1 F 13.5
20:00 258.2 F 20.1
22:00 268.9 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 29.1
2:00 271.9 F 29.3
3:00 272.1 F 28.9
4:00 272.2 F 29.5

Min 271.4 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 242.3 F 0.0
10:00 227.6 F 0.0
15:00 226.5 F 0.0
20:00 231.0 F 0.0
25:00 229.5 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:07
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:02:02
CYCLE COMPLETE

01/10/2025 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____