

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3710

Name: RITTER2  
Sterilizer ID: V2313228  
06/18/2024 10:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.0 F 0.0  
2:00 129.8 F 1.1  
4:00 137.8 F 3.6  
6:00 162.9 F 7.3  
8:00 186.4 F 11.8  
10:00 208.0 F 4.9  
12:00 224.2 F 9.7  
14:00 231.1 F 9.0  
16:00 238.8 F 11.7  
18:00 238.1 F 10.1  
20:00 251.3 F 16.6  
22:00 262.8 F 23.3  
24:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.7 F 29.9  
2:00 271.7 F 29.1  
3:00 272.4 F 30.0  
4:00 272.4 F 29.6

Min 271.5 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
5:00 205.7 F 0.0  
10:00 210.1 F 0.0  
15:00 207.5 F 0.0  
20:00 204.2 F 0.0  
25:00 204.8 F 0.0  
30:00 204.2 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 24:18  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:02:32  
CYCLE COMPLETE

06/18/2024 11:09 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_