
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000955
8:28 10/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:34
STERILIZING 3:34
136.0 C 220kPa 3:34
136.3 C 223kPa 4:04
136.3 C 223kPa 4:34
136.4 C 224kPa 5:04
136.4 C 221kPa 5:34
136.3 C 222kPa 6:04
136.3 C 223kPa 6:34
136.4 C 223kPa 7:04
136.4 C 223kPa 7:04
VENTING 7:04
AIR DRYING 7:56
CYCLE COMPLETE 8:15

Digital Signature #
9DB18DC4A25A6864
