
Midmark M11 - v1.0.0
Total Cycles: 3254

Name: RITTER 1
Sterilizer ID: V2178198
05/18/2023 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.9 F 0.0
2:00 155.0 F 0.0
4:00 162.9 F 0.0
6:00 188.7 F 4.1
8:00 205.4 F 8.4
10:00 225.0 F 13.0
12:00 226.8 F 8.0
14:00 239.8 F 13.1
16:00 239.5 F 10.6
18:00 236.5 F 8.8
20:00 251.3 F 15.8
22:00 263.2 F 22.6
24:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.5 F 28.9
2:00 271.8 F 28.2
3:00 271.6 F 28.4
4:00 272.4 F 28.8

Min 271.2 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 210.1 F 0.0
10:00 228.3 F 0.0
15:00 231.9 F 0.0
20:00 228.3 F 0.0
25:00 225.6 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 24:11
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:15
CYCLE COMPLETE

05/18/2023 10:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____