
Midmark M11 - v1.0.0
Total Cycles: 4668

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2024 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.2 F 0.0
2:00 160.6 F 2.4
4:00 162.4 F 5.6
6:00 182.4 F 9.3
8:00 211.7 F 11.0
10:00 216.4 F 6.8
12:00 226.9 F 10.9
14:00 228.9 F 6.8
16:00 240.7 F 12.0
18:00 238.7 F 10.0
20:00 251.4 F 16.1
22:00 262.2 F 22.3
24:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.7 F 29.4
2:00 271.8 F 28.7
3:00 272.7 F 29.4
4:00 271.9 F 28.8

Min 271.6 F 28.6
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 216.2 F 0.0
10:00 229.8 F 0.0
15:00 229.6 F 0.0
20:00 229.5 F 0.0
25:00 232.5 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:09
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:57
CYCLE COMPLETE

06/21/2024 12:16 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____