

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3859

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2023 10:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.4 F 0.0  
2:00 151.5 F 0.9  
4:00 165.9 F 4.0  
6:00 189.2 F 8.0  
8:00 209.8 F 12.6  
10:00 222.8 F 7.1  
12:00 234.1 F 12.0  
14:00 234.5 F 8.5  
16:00 242.8 F 11.4  
18:00 245.0 F 12.7  
20:00 257.0 F 18.9  
22:00 267.3 F 25.5  
24:00 272.4 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.1 F 28.7  
2:00 272.5 F 29.0  
3:00 272.8 F 29.4  
4:00 271.5 F 28.7

Min 271.5 F 28.6  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.7 F 0.0  
5:00 224.4 F 0.0  
10:00 196.5 F 0.0  
15:00 182.2 F 0.0  
20:00 178.3 F 0.0  
25:00 177.7 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 24:07  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:02:55  
CYCLE COMPLETE

10/04/2023 11:21 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_