
Midmark M11 - v1.0.0
Total Cycles: 2812

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2023 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.2 F 0.0
2:00 203.1 F 3.3
4:00 209.5 F 8.7
6:00 226.0 F 8.4
8:00 234.1 F 10.3
10:00 236.6 F 7.8
12:00 241.9 F 11.1
14:00 247.1 F 13.7
16:00 259.8 F 20.8
18:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.7 F 29.0
2:00 271.6 F 28.2
3:00 272.7 F 29.1
4:00 272.1 F 28.6

Min 271.5 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 225.1 F 0.0
10:00 235.3 F 0.0
15:00 232.0 F 0.0
20:00 227.2 F 0.0
25:00 225.6 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

01/24/2023 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____