
Midmark M11 - v1.0.0
Total Cycles: 900

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2021 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.1 F 0.0
2:00 135.5 F 1.0
4:00 160.0 F 3.4
6:00 186.3 F 7.0
8:00 203.3 F 11.3
10:00 220.3 F 5.1
12:00 226.6 F 9.9
14:00 233.1 F 8.8
16:00 240.6 F 11.9
18:00 238.4 F 10.0
20:00 251.8 F 16.5
22:00 263.5 F 23.1
24:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.1
2:00 271.7 F 28.7
3:00 272.4 F 29.1
4:00 271.6 F 28.6

Min 271.5 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 227.1 F 0.0
10:00 232.3 F 0.0
15:00 230.6 F 0.0
20:00 229.3 F 0.0
25:00 228.6 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 24:09
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:02:39
CYCLE COMPLETE

06/15/2021 02:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____