

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3293

Name: RITTER 1  
Sterilizer ID: V2178198  
05/29/2023 07:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.6 F 0.0  
2:00 75.4 F 0.0  
4:00 100.8 F 0.0  
6:00 132.9 F 0.0  
8:00 152.0 F 0.0  
10:00 175.9 F 3.0  
12:00 195.5 F 6.3  
14:00 208.3 F 9.7  
16:00 221.0 F 13.3  
18:00 224.5 F 6.7  
20:00 234.5 F 10.9  
22:00 232.7 F 7.3  
24:00 244.2 F 12.8  
26:00 241.6 F 10.9  
28:00 253.8 F 17.0  
30:00 264.2 F 23.2  
32:00 271.9 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.6 F 28.8  
2:00 271.7 F 28.1  
3:00 272.2 F 28.8  
4:00 272.3 F 28.6  
  
Min 271.4 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
5:00 203.8 F 0.0  
10:00 213.5 F 0.0  
15:00 216.2 F 0.0  
20:00 216.8 F 0.0  
25:00 220.1 F 0.0  
30:00 222.7 F 0.0  
  
Drying Complete

FILLING: 0:46  
HEATING: 32:10  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:10:11  
CYCLE COMPLETE

05/29/2023 08:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_