
Midmark M11 - v1.0.5
Total Cycles: 3954

Name: RITTER2
Sterilizer ID: V2313228
09/18/2024 03:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.7 F 0.0
2:00 132.2 F 1.4
4:00 142.2 F 4.1
6:00 167.8 F 7.5
8:00 189.6 F 11.1
10:00 217.8 F 5.4
12:00 222.2 F 7.6
14:00 230.9 F 11.8
16:00 232.2 F 8.3
18:00 244.6 F 13.0
20:00 243.5 F 12.4
22:00 254.8 F 18.2
24:00 265.0 F 24.3
26:00 271.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.8 F 30.0
2:00 271.8 F 29.3
3:00 271.7 F 29.8
4:00 272.7 F 29.8

Min 271.2 F 29.2
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 222.0 F 0.0
10:00 230.3 F 0.0
15:00 220.2 F 0.0
20:00 227.3 F 0.0
25:00 211.9 F 0.0
30:00 204.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 26:53
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:05:52
CYCLE COMPLETE

09/18/2024 05:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____