
Midmark M11 - v1.0.5
Total Cycles: 2975

Name: RITTER2
Sterilizer ID: V2313228
10/05/2023 07:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 72.5 F 0.0
4:00 82.3 F 0.4
6:00 108.9 F 2.0
8:00 140.8 F 4.6
10:00 166.3 F 7.9
12:00 189.1 F 11.7
14:00 214.7 F 4.1
16:00 223.7 F 7.6
18:00 228.9 F 11.5
20:00 228.6 F 6.9
22:00 240.4 F 11.9
24:00 237.7 F 9.6
26:00 249.5 F 15.4
28:00 259.9 F 21.3
30:00 270.0 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.4 F 29.4
2:00 272.9 F 29.7
3:00 272.5 F 30.1
4:00 272.0 F 29.6

Min 271.7 F 29.3
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.4 F 0.0
5:00 256.6 F 0.0
10:00 225.5 F 0.0
15:00 220.4 F 0.0
20:00 211.3 F 0.0
25:00 218.6 F 0.0
30:00 204.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:36
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:49
CYCLE COMPLETE

10/05/2023 08:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____