
Midmark M11 - v1.0.5
Total Cycles: 1667

Name: RITTER2
Sterilizer ID: V2313228
11/02/2022 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.7 F 0.0
2:00 151.7 F 1.8
4:00 164.0 F 4.7
6:00 184.7 F 8.8
8:00 213.1 F 10.9
10:00 219.1 F 8.5
12:00 230.7 F 12.1
14:00 236.9 F 10.4
16:00 235.1 F 8.2
18:00 248.1 F 14.8
20:00 259.7 F 21.3
22:00 270.1 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.5 F 29.8
2:00 271.2 F 29.1
3:00 272.7 F 30.0
4:00 272.0 F 29.5

Min 271.1 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 221.8 F 0.0
10:00 237.8 F 0.0
15:00 230.3 F 0.0
20:00 226.0 F 0.0
25:00 227.4 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:56
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

11/02/2022 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____