
Midmark M11 - v1.0.5
Total Cycles: 1423

Name: RITTER2
Sterilizer ID: V2313228
08/29/2022 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.9 F 0.0
2:00 184.7 F 1.7
4:00 180.1 F 4.8
6:00 206.7 F 10.1
8:00 221.0 F 5.7
10:00 232.4 F 10.8
12:00 233.8 F 8.0
14:00 242.1 F 11.5
16:00 245.1 F 13.3
18:00 257.1 F 19.8
20:00 268.0 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 271.8 F 29.5
2:00 272.6 F 29.7
3:00 271.6 F 29.0
4:00 271.8 F 29.6

Min 271.2 F 28.9
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 213.3 F 0.0
10:00 223.6 F 0.0
15:00 233.4 F 0.0
20:00 230.2 F 0.0
25:00 231.0 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:39
CYCLE COMPLETE

08/29/2022 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____