

-----  
Midmark M11 - v1.0.5  
Total Cycles: 100

Name: RITTER2  
Sterilizer ID: V2313228  
06/15/2021 07:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.7 F 0.0  
2:00 130.6 F 1.1  
4:00 134.3 F 3.4  
6:00 153.7 F 6.9  
8:00 181.2 F 11.2  
10:00 210.3 F 4.5  
12:00 217.0 F 9.1  
14:00 231.6 F 11.1  
16:00 235.7 F 11.2  
18:00 237.1 F 9.6  
20:00 249.6 F 15.7  
22:00 260.6 F 22.1  
24:00 271.4 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.7  
1:00 272.7 F 30.0  
2:00 272.2 F 30.2  
3:00 271.3 F 29.8  
4:00 272.0 F 29.8

Min 271.1 F 29.6  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.3 F 0.0  
5:00 222.6 F 0.0  
10:00 197.5 F 0.0  
15:00 196.1 F 0.0  
20:00 195.5 F 0.0  
25:00 195.1 F 0.0  
30:00 194.9 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 25:27  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:03:49  
CYCLE COMPLETE

06/15/2021 08:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_