
Midmark M11 - v1.0.5
Total Cycles: 2674

Name: RITTER2
Sterilizer ID: V2313228
07/20/2023 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 215.3 F 0.0
2:00 202.8 F 3.1
4:00 206.6 F 8.6
6:00 222.9 F 9.9
8:00 231.0 F 10.2
10:00 232.6 F 8.0
12:00 241.6 F 11.7
14:00 245.2 F 13.3
16:00 258.1 F 20.2
18:00 269.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.2 F 29.4
2:00 271.3 F 29.2
3:00 272.4 F 29.4
4:00 271.1 F 28.9

Min 271.1 F 28.8
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 213.9 F 0.0
10:00 232.0 F 0.0
15:00 237.4 F 0.0
20:00 231.2 F 0.0
25:00 221.4 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

07/20/2023 11:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____