

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.3uS / 0.8ppm  
CYCLE NUMBER 000597  
12:31 25/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:08  
STERILIZING 4:26  
136.0 C 221kPa 4:26  
136.4 C 230kPa 4:56  
136.4 C 228kPa 5:26  
136.2 C 229kPa 5:56  
136.4 C 232kPa 6:26  
136.2 C 229kPa 6:56  
136.4 C 231kPa 7:26  
136.3 C 231kPa 7:56  
136.3 C 231kPa 7:56  
VENTING 7:56  
AIR DRYING 8:49  
CYCLE COMPLETE 23:49  
-----

Digital Signature #  
C58E8EE47686B397  
-----