
Midmark M11 - v1.0.0
Total Cycles: 3494

Name: RITTER 1
Sterilizer ID: V2178198
07/06/2023 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.8 F 0.0
2:00 173.1 F 0.2
4:00 185.7 F 3.5
6:00 206.3 F 8.5
8:00 221.5 F 10.7
10:00 229.1 F 9.1
12:00 236.9 F 9.9
14:00 242.2 F 12.1
16:00 240.2 F 10.5
18:00 253.3 F 17.1
20:00 264.8 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.9
2:00 271.6 F 28.6
3:00 272.4 F 28.9
4:00 271.4 F 28.4

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 207.8 F 0.0
10:00 215.9 F 0.0
15:00 215.3 F 0.0
20:00 219.6 F 0.0
25:00 215.7 F 0.0
30:00 213.8 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:52
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

07/06/2023 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____