
Midmark M11 - v1.0.5
Total Cycles: 2296

Name: RITTER2
Sterilizer ID: V2313228
04/25/2023 07:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.7 F 0.0
2:00 80.0 F 0.0
4:00 86.4 F 0.0
6:00 114.2 F 0.0
8:00 147.7 F 1.5
10:00 175.4 F 4.6
12:00 201.2 F 8.2
14:00 208.2 F 12.7
16:00 221.3 F 6.3
18:00 231.6 F 10.6
20:00 232.2 F 8.1
22:00 242.5 F 13.0
24:00 242.0 F 11.6
26:00 254.3 F 18.0
28:00 265.4 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.3 F 29.5
2:00 271.4 F 28.7
3:00 272.4 F 29.7
4:00 272.0 F 29.0

Min 271.1 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.2 F 0.0
5:00 215.7 F 0.0
10:00 225.4 F 0.0
15:00 214.4 F 0.0
20:00 208.0 F 0.0
25:00 205.8 F 0.0
30:00 204.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:49
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:07:48
CYCLE COMPLETE

04/25/2023 08:19 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____