
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001677
13:47 01/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:00
STERILIZING 2:53
136.4 C 220kPa 2:53
136.4 C 225kPa 3:23
136.3 C 224kPa 3:53
136.1 C 223kPa 4:23
136.4 C 223kPa 4:53
136.4 C 224kPa 5:23
136.2 C 222kPa 5:53
136.4 C 225kPa 6:23
136.4 C 225kPa 6:23
VENTING 6:23
AIR DRYING 7:12
CYCLE COMPLETE 7:14

Digital Signature #
5176A1F35A046B0E
