

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
2.4uS / 1.5ppm  
CYCLE NUMBER 003743  
15:21 31/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 0:54  
PRESSURIZING 1:00  
STERILIZING 2:18  
136.5 C 223kPa 2:18  
136.3 C 224kPa 2:48  
136.3 C 224kPa 3:18  
136.2 C 224kPa 3:48  
136.4 C 225kPa 4:18  
136.3 C 223kPa 4:48  
136.2 C 220kPa 5:18  
136.4 C 224kPa 5:48  
136.4 C 224kPa 5:48  
VENTING 5:48  
AIR DRYING 6:42  
CYCLE COMPLETE 7:22

-----  
Digital Signature #  
8D2D1A9787D9FB54  
-----