
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001677
10:34 10/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:04
STERILIZING 4:02
136.4 C 222kPa 4:02
136.3 C 225kPa 4:32
136.4 C 224kPa 5:02
136.4 C 224kPa 5:32
136.1 C 223kPa 6:02
136.4 C 225kPa 6:32
136.3 C 223kPa 7:02
136.4 C 221kPa 7:32
136.4 C 221kPa 7:32
VENTING 7:32
AIR DRYING 8:21
CYCLE COMPLETE 10:20

Digital Signature #
FDD2301AA6A6F53A
