

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5356

Name: RITTER 1  
Sterilizer ID: V2178198  
02/21/2025 03:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.2 F 0.0  
2:00 96.2 F 1.0  
4:00 126.8 F 3.4  
6:00 143.5 F 6.0  
8:00 193.6 F 9.1  
10:00 204.3 F 13.2  
12:00 214.3 F 3.9  
14:00 220.2 F 7.2  
16:00 232.2 F 11.0  
18:00 231.3 F 6.8  
20:00 241.6 F 11.5  
22:00 237.6 F 9.1  
24:00 249.9 F 14.9  
26:00 260.2 F 20.7  
28:00 269.8 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.7 F 29.2  
2:00 271.8 F 28.5  
3:00 272.7 F 29.3  
4:00 272.4 F 28.8  
  
Min 271.6 F 28.5  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.6 F 0.0  
5:00 197.9 F 0.0  
10:00 210.4 F 0.0  
15:00 206.5 F 0.0  
20:00 197.1 F 0.0  
25:00 197.9 F 0.0  
30:00 202.1 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:52  
STERILIZING: 4:00  
VENTING: 3:59  
DRYING: 30:00  
TOTAL CYCLE: 01:08:42  
CYCLE COMPLETE

02/21/2025 04:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_