
Midmark M11 - v1.0.5
Total Cycles: 3339

Name: RITTER2
Sterilizer ID: V2313228
01/31/2024 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.6 F 0.0
2:00 142.8 F 0.8
4:00 149.2 F 2.9
6:00 178.9 F 6.6
8:00 205.4 F 11.0
10:00 221.2 F 5.1
12:00 227.4 F 10.1
14:00 232.0 F 8.1
16:00 242.3 F 12.5
18:00 240.9 F 11.1
20:00 253.4 F 17.4
22:00 264.6 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.5 F 29.8
2:00 272.5 F 29.7
3:00 272.6 F 29.9
4:00 271.9 F 29.9

Min 270.6 F 28.6
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.9 F 0.0
5:00 238.9 F 0.0
10:00 195.4 F 0.0
15:00 181.9 F 0.0
20:00 178.4 F 0.0
25:00 176.4 F 0.0
30:00 175.4 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:49
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:02:15
CYCLE COMPLETE

01/31/2024 11:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____