
Midmark M11 - v1.0.0
Total Cycles: 2949

Name: RITTER 1
Sterilizer ID: V2178198
02/27/2023 03:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.5 F 0.0
2:00 145.8 F 1.7
4:00 172.7 F 5.0
6:00 198.7 F 9.6
8:00 216.9 F 6.7
10:00 227.6 F 9.2
12:00 234.2 F 10.0
14:00 240.2 F 11.6
16:00 238.8 F 10.0
18:00 252.8 F 17.0
20:00 264.9 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.6
3:00 272.5 F 29.1
4:00 271.4 F 28.4

Min 271.3 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 223.2 F 0.0
10:00 223.4 F 0.0
15:00 221.9 F 0.0
20:00 225.6 F 0.0
25:00 229.8 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

02/27/2023 04:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____