
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005984
10:58 17/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 2:59
STERILIZING 3:58
136.4 C 221kPa 3:58
136.4 C 225kPa 4:28
136.3 C 223kPa 4:58
136.4 C 225kPa 5:28
136.3 C 224kPa 5:58
136.4 C 224kPa 6:28
136.4 C 224kPa 6:58
136.2 C 221kPa 7:28
136.2 C 221kPa 7:28
VENTING 7:28
AIR DRYING 8:23
CYCLE COMPLETE 9:01

Digital Signature #
D80670FC783AF358
