

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1735

Name: RITTER 1  
Sterilizer ID: V2178198  
03/23/2022 12:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.5 F 0.0  
2:00 136.7 F 0.3  
4:00 155.6 F 2.5  
6:00 182.5 F 5.8  
8:00 201.9 F 10.3  
10:00 221.1 F 5.3  
12:00 229.9 F 9.5  
14:00 235.8 F 8.8  
16:00 242.7 F 12.2  
18:00 240.9 F 10.6  
20:00 254.4 F 17.5  
22:00 266.1 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.2  
1:00 272.4 F 28.9  
2:00 271.3 F 28.4  
3:00 272.1 F 28.5  
4:00 271.8 F 28.8

Min 271.1 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.2 F 0.0  
5:00 229.2 F 0.0  
10:00 202.0 F 0.0  
15:00 205.5 F 0.0  
20:00 233.5 F 0.0  
25:00 234.3 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 23:27  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:01:32  
CYCLE COMPLETE

03/23/2022 01:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_