
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000828
15:38 30/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 5:34
STERILIZING 6:57
136.0 C 220kPa 6:57
136.3 C 221kPa 7:27
136.3 C 221kPa 7:57
136.2 C 221kPa 8:27
136.3 C 222kPa 8:57
136.3 C 220kPa 9:27
136.2 C 220kPa 9:57
136.4 C 222kPa 10:27
136.4 C 222kPa 10:27
VENTING 10:27
AIR DRYING 11:28
CYCLE COMPLETE 17:01

Digital Signature #
0955701437F19203
