
Midmark M11 - v1.0.5
Total Cycles: 978

Name: RITTER2
Sterilizer ID: V2313228
04/12/2022 10:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.9 F 0.0
2:00 160.7 F 0.1
4:00 184.4 F 1.1
6:00 213.4 F 5.9
8:00 221.6 F 11.1
10:00 228.2 F 6.9
12:00 240.6 F 12.3
14:00 239.4 F 10.5
16:00 237.3 F 9.4
18:00 251.5 F 16.2
20:00 263.1 F 23.0
22:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.7 F 29.3
2:00 271.6 F 28.5
3:00 272.2 F 29.4
4:00 272.5 F 29.0

Min 271.2 F 28.4
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.9 F 0.0
5:00 235.6 F 0.0
10:00 225.5 F 0.0
15:00 227.7 F 0.0
20:00 226.8 F 0.0
25:00 223.5 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:12
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:32
CYCLE COMPLETE

04/12/2022 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____