
Midmark M11 - v1.0.0
Total Cycles: 5413

Name: RITTER 1
Sterilizer ID: V2178198
03/13/2025 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.2 F 0.0
2:00 110.4 F 1.1
4:00 118.8 F 3.5
6:00 158.0 F 6.3
8:00 194.4 F 9.5
10:00 219.5 F 8.2
12:00 215.1 F 4.9
14:00 224.8 F 8.7
16:00 234.9 F 12.9
18:00 234.3 F 8.9
20:00 244.2 F 12.3
22:00 242.7 F 11.9
24:00 254.0 F 17.5
26:00 264.0 F 23.4
28:00 272.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.5 F 29.6
2:00 272.7 F 29.2
3:00 271.6 F 28.7
4:00 271.9 F 29.3

Min 271.4 F 28.7
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 190.2 F 0.0
10:00 206.8 F 0.0
15:00 175.6 F 0.0
20:00 185.2 F 0.0
25:00 190.8 F 0.0
30:00 199.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:08
STERILIZING: 4:00
VENTING: 4:03
DRYING: 30:00
TOTAL CYCLE: 01:08:02
CYCLE COMPLETE

03/13/2025 03:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____