

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5463

Name: RITTER 1  
Sterilizer ID: V2178198  
03/31/2025 10:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.0 F 0.0  
2:00 138.5 F 1.8  
4:00 140.0 F 4.5  
6:00 161.1 F 7.7  
8:00 184.9 F 11.3  
10:00 209.3 F 3.4  
12:00 215.0 F 7.4  
14:00 226.5 F 11.5  
16:00 229.5 F 7.0  
18:00 241.4 F 12.1  
20:00 238.1 F 9.9  
22:00 249.9 F 15.5  
24:00 260.3 F 21.3  
26:00 269.9 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.5 F 29.6  
2:00 271.6 F 29.0  
3:00 272.5 F 29.7  
4:00 272.2 F 29.2

Min 271.5 F 29.0  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
5:00 217.2 F 0.0  
10:00 211.9 F 0.0  
15:00 229.7 F 0.0  
20:00 209.1 F 0.0  
25:00 205.7 F 0.0  
30:00 200.7 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 27:49  
STERILIZING: 4:00  
VENTING: 3:59  
DRYING: 30:00  
TOTAL CYCLE: 01:06:38  
CYCLE COMPLETE

03/31/2025 11:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_