
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008374
8:45 05/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:13
STERILIZING 4:17
136.4 C 222kPa 4:17
136.1 C 219kPa 4:47
136.4 C 220kPa 5:17
136.1 C 222kPa 5:47
136.1 C 220kPa 6:17
136.3 C 221kPa 6:47
135.8 C 219kPa 7:17
136.4 C 219kPa 7:47
136.4 C 219kPa 7:47
VENTING 7:48
AIR DRYING 8:40
CYCLE COMPLETE 8:42

Digital Signature #
E02689BB2A85C4B3
