
Midmark M11 - v1.0.0
Total Cycles: 2664

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2022 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.6 F 0.0
2:00 88.0 F 0.0
4:00 124.3 F 0.0
6:00 128.3 F 0.8
8:00 155.0 F 2.4
10:00 176.2 F 4.8
12:00 193.3 F 8.0
14:00 208.3 F 11.4
16:00 215.7 F 4.4
18:00 227.1 F 8.1
20:00 237.3 F 12.3
22:00 235.2 F 8.7
24:00 241.7 F 10.9
26:00 245.3 F 12.7
28:00 257.0 F 19.0
30:00 267.6 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.8
2:00 271.4 F 28.1
3:00 272.7 F 28.8
4:00 271.6 F 28.1

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 231.0 F 0.0
10:00 202.6 F 0.0
15:00 191.9 F 0.0
20:00 196.3 F 0.0
25:00 199.5 F 0.0
30:00 203.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:24
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:32
CYCLE COMPLETE

12/13/2022 09:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____