
Midmark M11 - v1.0.0
Total Cycles: 5324

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2025 02:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.6 F 0.0
2:00 102.4 F 0.9
4:00 116.7 F 3.3
6:00 147.6 F 6.1
8:00 173.1 F 9.0
10:00 191.1 F 12.3
12:00 211.5 F 3.0
14:00 215.9 F 6.4
16:00 226.9 F 10.0
18:00 235.8 F 10.1
20:00 238.1 F 9.8
22:00 239.8 F 10.2
24:00 245.0 F 12.5
26:00 255.5 F 18.1
28:00 265.2 F 24.0
30:00 271.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.1 F 29.2
2:00 272.4 F 29.2
3:00 272.4 F 28.9
4:00 271.5 F 28.5

Min 271.3 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.3 F 0.0
5:00 220.6 F 0.0
10:00 226.3 F 0.0
15:00 248.4 F 0.0
20:00 228.7 F 0.0
25:00 213.4 F 0.0
30:00 195.7 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:07
STERILIZING: 4:00
VENTING: 3:59
DRYING: 30:00
TOTAL CYCLE: 01:09:51
CYCLE COMPLETE

02/07/2025 03:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____