

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2451

Name: RITTER2  
Sterilizer ID: V2313228  
06/01/2023 10:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
2:00 196.3 F 3.8  
4:00 199.7 F 9.1  
6:00 223.5 F 8.0  
8:00 233.4 F 10.8  
10:00 233.8 F 8.3  
12:00 240.7 F 10.5  
14:00 247.2 F 14.2  
16:00 260.0 F 21.4  
18:00 271.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.6 F 29.6  
2:00 271.3 F 28.9  
3:00 272.3 F 29.8  
4:00 272.3 F 29.4  
  
Min 271.2 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 218.3 F 0.0  
10:00 240.5 F 0.0  
15:00 238.2 F 0.0  
20:00 225.6 F 0.0  
25:00 222.2 F 0.0  
30:00 221.5 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 18:40  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:43  
CYCLE COMPLETE

06/01/2023 11:34 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_