
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001030
11:24 08/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:39
STERILIZING 3:50
136.4 C 225kPa 3:50
136.3 C 224kPa 4:20
136.1 C 226kPa 4:50
136.3 C 224kPa 5:20
136.4 C 225kPa 5:50
136.1 C 226kPa 6:20
136.1 C 223kPa 6:50
136.4 C 225kPa 7:20
136.4 C 225kPa 7:20
VENTING 7:20
AIR DRYING 8:11
CYCLE COMPLETE 23:11

Digital Signature #
B02B90150381BABE
