
Midmark M11 - v1.0.5
Total Cycles: 986

Name: RITTER2
Sterilizer ID: V2313228
04/13/2022 03:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.1 F 0.0
2:00 185.5 F 0.8
4:00 198.9 F 5.8
6:00 217.5 F 11.5
8:00 227.3 F 8.0
10:00 239.5 F 11.7
12:00 243.0 F 12.9
14:00 243.9 F 12.7
16:00 257.7 F 20.2
18:00 269.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.6 F 29.6
2:00 271.8 F 29.3
3:00 272.6 F 29.6
4:00 271.6 F 29.0

Min 271.6 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 206.1 F 0.0
10:00 221.4 F 0.0
15:00 229.1 F 0.0
20:00 231.8 F 0.0
25:00 224.8 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:35
CYCLE COMPLETE

04/13/2022 04:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____